

PERFORMANCE CHARACTERISTICS (continued)**Sample Recovery**

High and low concentrations of purified human IgG were spiked into each of 3 serum samples. Observed assay values compared to expected values ranged from 101 to 135%, indicating accurate quantification of IgG in human serum.

Sample	Expected ng/ml	Observed ng/ml	Observed/ Expected
High IgG Spike		85.4	
+ Human D, 48.2 ng/ml	134	136	102 %
+ Human E, 80.8 ng/ml	166	178	107 %
+ Human F, 36.4 ng/ml	122	146	120 %
Low IgG Spike		27.6	
+ Human D, 48.2 ng/ml	76	78	102 %
+ Human E, 80.8 ng/ml	108	110	101 %
+ Human F, 36.4 ng/ml	64	86	135 %

ELISA Kit Components	Amount	Part No.
Anti-Human IgG Microwell Strip Plate	8-well strips (12)	1751
Human IgG Positive Control	0.65 ml	1752
Human IgG Standard 10 ng/ml	0.65 ml	1753B
Human IgG Standard 25 ng/ml	0.65 ml	1753C
Human IgG Standard 50 ng/ml	0.65 ml	1753D
Human IgG Standard 100 ng/ml	0.65 ml	1753E
Human IgG Standard 200 ng/ml	0.65 ml	1753F
Anti-Human IgG HRP Conjugate (100X)	0.15 ml	1754
Sample Diluent Concentrate (20X)	10 ml	SD-20T
Wash Solution Concentrate (100X)	10 ml	WB-100
TMB Substrate	12 ml	80091
Stop Solution	12 ml	80101
Product Manual	1 ea	M-1750

Related Items

Catalog#	ProdDescription
1750	Human IgG ELISA Kit, 96 tests, Quantitative
1750-RS	Human IgG Reference Serum for ELISA (~500 ng/ml)
1760	Human IgM ELISA Kit, 96 tests, Quantitative
1760-RS	Human IgM Reference Serum for ELISA (~500 ng/ml)
1780	Human IgA ELISA Kit, 96 tests, Quantitative
1800	Human IgE ELISA Kit, 96 tests, Quantitative

Instruction Manual No. M-1750

Human IgG

ELISA Kit Cat. No. 1750

For Quantitative Determination of Human Immunoglobulin G serum, plasma or other biological Fluids

For in vitro research use only (RUO), not for therapeutic or diagnostic use.



**ALPHA DIAGNOSTIC
INTERNATIONAL**

6203 Woodlake Center Drive • San Antonio • Texas 78244 • USA.

Phone (210) 561-9515 • Fax (210) 561-9544


Toll Free (800) 786-5777

Email: service@4adi.com
www.4adi.com

INTENDED USE

The Human IgG ELISA Kit is an in vitro immunoassay for the quantification of IgG circulating in serum, plasma or in other appropriately qualified samples from tissue fluids (e.g., saliva, mucosa), or in cultures of human cells. For in vitro research use only (RUO), not for therapeutic or diagnostic use.

RESEARCH USE OF THE TEST



Immunoassays using heavy-chain specific antibodies provide for selective, sensitive quantification of human immunoglobulins IgG, IgA and IgM, as found circulating in blood or as present in other body fluids, including saliva, milk/colostrum, ascites, tears and mucosa of linings of the gut, respiratory or urogenital tracts.

Levels of total IgG, IgA and/or IgM can reveal health status or results of experimental or pathological conditions (e.g., hypo- or hypergammaglobulinemia or acute or chronic infection). Also, measurements of specific antibody levels, in antigen-specific assays, are

often best interpreted relative to values of total IgG, IgA, and IgM in the sample and/or individual.

The quantitative immunoassays measure human IgG, IgA and IgM with high sensitivity; this allows for sufficient dilution of the sample to avoid sample matrix interference that may occur with any of the above specimen types. Also, each assay is Ig class specific, such that all IgG or IgA subclasses are reliably quantified in essentially any specimen, freshly obtained and/or suitable stored. Expected performance of each kit relative to precision, recovery and linearity of dilution is presented as guidance for use and experimental design.

PRINCIPLE OF THE TEST

The Human IgG ELISA kit is based on the binding of human IgG in samples to two antibodies, one immobilized on the microtiter wells, and the other conjugated to horseradish peroxidase (HRP) enzyme. After a washing step, chromogenic substrate is added and color is developed by the enzymatic reaction of HRP on the TMB substrate, which is directly proportional to the amount of IgG present in the sample. Stopping Solution is added to terminate the reaction, and absorbance at 450nm is then measured using an ELISA microtiter well reader. The concentration of IgG in samples and control is calculated from a curve of standards containing known concentrations of IgG.

STORAGE AND STABILITY

The microtiter well plate and all other reagents, if unopened, are stable at 2-8°C until the expiration date printed on the label. Stabilities of the working solutions are indicated under Reagent Preparation.

PERFORMANCE CHARACTERISTICS

Specificity

The antibodies used in this kit have been shown by immunoelectrophoresis and ELISA to react specifically with IgG, and have essentially no reactivity with IgM, IgA, IgE or any other human serum proteins.

Serum from the following species showed significant reactivity at 1:1000 dilution: monkey, rat, donkey, dog, cat, rabbit. The following showed no significant reactivity: hamster, guinea pig, sheep, goat, bovine, chicken; also 10% neonatal bovine serum.

Normal Range

Assay of IgG in stored sera from twenty (20) individuals ranged from 3.5 to 15 mg/ml (median = 10 mg/ml). Each laboratory should determine expected values of its own testing population.

Precision

Samples containing low, medium and high concentrations of IgG, representing 3 different sera, were assayed multiple times in the same assay (n=10) to provide within-assay precision, and as duplicates in multiple assays (n=5) to obtain between-assay reproducibility. Coefficient of variations were calculated for the concentrations using a point-to-point curve-fitting program.

IgG concentrations were measured with very good within-assay (3.9 to 6.8 %CV) and good between-assay (6.4 to 13.2 %CV) reproducibility.

Sample	IgG ng/ml	Intra-assay %CV	Inter-assay %CV
Human Serum A	36.0	3.9	6.4
Human Serum B	61.2	4.0	13.2
Human Serum C	143	6.8	6.8

Linearity of Dilution

Three (3) individual stored sera and a purified IgG preparation were diluted to 2 levels for testing, and concordance of the assay values were compared. The mean recovery ranged from 93 to 99%, demonstrating linear dilution and equivalent quantification across the standard range.

Sample	Dilution	Assay Value ng/ml	Serum Value mg/ml	Concordance
Human Serum D	1:40k	192	7.70	94 %
	1:320k	21.5	6.86	
Human Serum E	1:40k	115	4.61	95 %
	1:320k	12.9	4.13	
Human Serum F	1:25k	151	3.77	93 %
	1:200k	16.3	3.26	
Purified IgG	1:5k	190	0.95	99 %
	1:40k	24.4	0.97	

Continued on Page 7.

CALCULATION OF RESULTS


- The results may be calculated using any immunoassay software package. The four-parameter curve-fit is recommended. If software is not available, IgG concentrations may be determined as follows:
- Calculate the mean OD of duplicate samples.
- On graph paper plot the mean OD of the standards (y-axis) against the concentration (ng/ml) of IgG (x-axis). Draw the best fit curve through these points to construct the standard curve. A point-to-point construction is most common and reliable.
- The IgG concentrations in unknown samples and controls can be determined by interpolation from the standard curve.
- Multiply the values obtained for the samples by the dilution factor of each sample.
- Samples producing signals higher than the 200 ng/ml standard should be further diluted and re-assayed.

TYPICAL RESULTS

The following data are for illustration purposes only. A complete standard curve should be run in every assay to determine sample values.

Wells	Standards, Control & Samples	A450 nm	IgG ng/ml
A1, A2	Negative Diluent Control	0.05	0
B1, B2	10 ng/ml Standard	0.27	10
C1, C2	25 ng/ml Standard	0.68	25
D1, D2	50 ng/ml Standard	1.17	50
E1, E2	100 ng/ml Standard	1.71	100
F1, F2	200 ng/ml Standard	2.32	200
G1, G2	Positive Control [Value: 42 - 78 ng/ml]	1.32	64
H1, H2	Sample [Diluted 1:50k] Calculated: 50k-fold dilution x 155 ng/ml = 7.75 mg/ml in serum	2.04	155

A typical assay Standard Curve (do not use for calculating sample values)



/bill-elisa_graph-1

KIT CONTENTS

To Be Reconstituted: Store as indicated.

Component	Instructions for Use
Sample Diluent Concentrate (20x) Cat. No. SD-20T, 10ml	Dilute the entire volume, 10ml + 190ml with distilled or deionized water into a clean stock bottle. Label as Working Sample Diluent and store at 2-8°C until the kit lot expires or is used up.
Wash Solution Concentrate (100x) Cat. No. WB-100, 10ml	Dilute the entire volume 10ml + 990ml with distilled or deionized water into a clean stock bottle. Label as Working Wash Solution and store at RT until kit is used entirely.
Anti-Human IgG - HRP Conjugate Concentrate (100x) Part No. 1764, 0.15ml	Peroxidase conjugated anti-human IgG in buffer with protein, detergents and ProClin 300 as stabilizers. Dilute fresh as needed; 10ul of concentrate to 1ml of Working Sample Diluent is sufficient for 1 8-well strip. Use within the working day and discard. Return concentrate to 2-8°C storage.

Ready For Use: Store as indicated on labels.

Component	Part No.	Amt	Contents
Anti-Human IgG Microwell Strip Plate	1761	8-well strips (12)	Coated with purified anti-human IgG antibodies.
Human IgG Standards			
10 ng/ml	1763B	0.65 ml	Five (5) vials, each containing human serum with calibrated IgG concentrations; diluted in buffer with protein, detergents and ProClin 300 as stabilizers.
25 ng/ml	1763C	0.65 ml	
50 ng/ml	1763D	0.65 ml	
100 ng/ml	1763E	0.65 ml	
200 ng/ml	1763F	0.65 ml	
Positive Control [IgG] range on label	1762	0.65 ml	Human serum with stated IgG concentration range; diluted in buffer with protein, detergents and ProClin 300 as stabilizers.
TMB Substrate	80091	12 ml	Chromogenic substrate for HRP containing TMB and peroxide.
Stop Solution	80101	12 ml	Diluted sulfuric acid.

Materials Required But Not Provided:

- Pipettors and pipettes that deliver 100ul and 1-10ml. A multi-channel pipetter is recommended.
- Disposable glass or plastic 5-15ml tubes for diluting samples and anti-human IgG-HRP Concentrate.
- Graduated cylinder to dilute Wash Concentrate and Sample Diluent concentrate; 200ml to 1L.
- Stock bottle to store diluted Wash Solution; 200ml to 1L.
- Distilled or deionized water to dilute reagent concentrates.
- Microwell plate reader at 450 nm wavelength.

SPECIMEN COLLECTION AND HANDLING

Collect blood by venipuncture, allow clotting, and separate the serum by centrifugation at room temperature. Do not heat inactivate the serum. If sera are not assayed immediately, stored refrigerated for up to 2 weeks, or frozen for long-term storage. Avoid freeze-thaw cycles.

The use of plasma has not been investigated, but should be a suitable specimen for assay.

PRECAUTIONS AND SAFETY INSTRUCTIONS

Human serum and other bodily fluids may contain infectious material. Always wear gloves when handling human samples, including the standards and controls, and dispose of these samples and containers as biohazard waste.

Standards, Controls, Sample Diluent, and Anti-human IgG-HRP contain Proclin 300 (0.05%, v/v). Stop Solution contains 1% sulfuric acid. Follow good laboratory practices, and avoid ingestion or contact of any reagent with skin, eyes or mucous membranes. All reagents may be disposed of down a drain with copious amounts of water.

Applicable **MSDS**, if not already on file, for the following reagents can be obtained from ADI web site. TMB (substrate), H₂SO₄ (stop solution), and Proclin-300 (0.1% v/v in standards, sample diluent and HRP-conjugates).

http://4adi.com/commerce/info/showpage.jsp?page_id=1060&category_id=2430&visit=10

QUALITY CONTROL

Reagents Accurate and reproducible assay results rely on proper storage, handling and control of reagent and sample temperature. Store all reagents as indicated, and warm to room temperature only those to be used in the assay. Shelf-life of the critical reagents and samples will diminish with extended exposure to non-refrigeration, resulting in inaccurate assay results. All solutions should be clear. Cloudiness or particulates are indications of reagent contamination or instability and may interfere with proper performance of the assay. Do not use.

Sample Controls A Positive Serum Control is provided with the kit, assigned with an IgG concentration value range. Recovery in this range is an indicator of proper assay performance. Each lab should also assay internal control samples, which represent the lab's expected sample population and that are maintained stabilized. A Negative Diluent Control should also be run.

ASSAY PROCEDURE

Bring all reagents to room temperature (18-30° C) equilibration (at least 30 minutes).

DILUTE Serum Samples in Working Sample Diluent. Dilutions of 50 to 200k-fold are appropriate for most normal human sera. For accuracy, three dilution steps are recommended, as follows:

- 1) 20ul serum + 380ul diluent = [1:20],
- 2) 20ul [1:20] + 980ul diluent = [1:1k],
- 3) 20ul [1:1k] + 980ul diluent = **1:50k**

DO NOT dilute the Standards or Control.

ALL STEPS ARE PERFORMED AT ROOM TEMPERATURE. After each reagent addition, gently tap the plate to mix the well contents prior to beginning incubation.

1. Set-up

- Determine the number of wells for the assay run. Duplicates are recommended, including 10 Standard wells and 2 wells for each sample and control to be assayed.
- Remove the appropriate number of microwell strips from the pouch and return unused strips to the pouch. Reseal the pouch and store refrigerated.
- Add 200-300ul Working Wash Solution to each well and let stand for 5 to 30 minutes before sample addition.
- Aspirate the liquid and pat dry on a paper towel.

2. 1st Incubation

[100ul - 60min; 4 washes]

- Add 100ul of standards, samples and controls each to pre-determined wells.
- Tap the plate gently to mix reagents and incubate for 60 minutes.
- Wash wells 4 times and pat dry on fresh paper towels. As an alternative, an automatic plate washer is recommended. Improper washes may lead to falsely elevated signals and poor reproducibility.

3. 2nd Incubation

[100ul - 30min; 5 washes]

- Add 100ul of diluted Anti-human IgG-HRP Conjugate to each well.
- Incubate for 30 minutes.
- Wash wells 5 times as in step 2.

4. Substrate Incubation

[100ul - 15min]

- Add 100ul TMB Substrate to each well. The liquid in the wells will begin to turn blue.
 - Incubate for 15 minutes in the dark, e.g., place in a drawer or closet.
- Note: If your microplate reader does not register optical density (OD) above 2.0, incubate for less time, assuring the top standard does not surpass 2 OD.

5. Stop Step

[Stop: 100ul]

- Add 100ul of Stop Solution to each well.
- Tap gently to mix. The enzyme reaction will stop; liquid in the wells will turn yellow.

6. Absorbance Reading

- Use any commercially available microplate reader capable of reading at 450nm wavelength. Use a program suitable for obtaining OD readings, and data calculations if available.
- Read absorbance of the entire plate at 450nm using a single wavelength within 30 minutes after Stop Solution addition. If available, program to subtract OD at 630nm to normalize well background.