

## INTENDED USE


The Mouse anti-ssDNA Total Ig's ELISA Kit is an immunoassay suitable for quantifying or titrating total antibody activity (IgG, IgA and IgM) specific for ssDNA in serum or plasma. Other biological fluids, including tissue culture medium, may be validated for use.

## GENERAL INFORMATION

Antibodies reactive with autologous nuclear components, such as DNA and histones, can represent an autoimmune basis for pathological conditions such as systemic lupus erythematosus (SLE) in humans, and in mice homozygous for the lymphoproliferation spontaneous mutation (Fas<sup>lpr</sup>), a systemic autoimmunity with massive lymphadenopathy associated with proliferation of aberrant T cells, arthritis and immune complex glomerulonephritis. These conditions include elevated levels of anti-dsDNA and other anti-nuclear antibodies (ANA) which often increase as the animal ages. Also, the expanded use in the drug industry of biological modifiers has been associated with production of autoantibodies, of which mice, and possibly also other hosts such as humans and monkeys, are susceptible. A prototype disease in mice is lupus caused by the drug minocycline, with elevated anti-dsDNA among other autoantibodies and pathological conditions.

Recent investigations have focused on the role of innate immune mechanisms, including Toll-like receptors (TLRs) and TH2 immunity, responding to the damage-associated molecular patterns of dying cells, as underlying cause of the anti-dsDNA type autoimmunity; these may be induced by drugs, including vaccines and adjuvants, with aging, or with other conditions.

## PRINCIPLE OF THE TEST



The Mouse anti-ssDNA Total Ig ELISA kit is based on the binding of mouse anti-ssDNA IgG, IgA and IgM in samples to ssDNA immobilized on the microwells, and total anti-ssDNA antibody is detected by anti-mouse IgG+IgA+IgM-HRP conjugate. After a washing step, chromogenic substrate (TMB) is added and color (blue) is developed, which is directly proportional to the amount of anti-ssDNA Ig present in the sample. Stopping Solution is added to terminate the reaction (converts blue to yellow), and A450nm is then measured using an ELISA reader. The activity of total mouse antibody in samples is calculated relative calibrators.

## PRODUCT SPECIFICATIONS

### Specificity

Purified ssDNA is used to coat the microwells; thus the assay is specific for antibodies directed to ssDNA. The anti-Mouse IgG+IgA+IgM (H+L) HRP conjugate reacts with mouse IgG, IgA and IgM class antibodies that bind to ssDNA on the plate. IgE antibody would not be measured above background signals.

### Assay Sensitivity

The ssDNA coating level, HRP conjugate concentration and Low NSB Sample Diluent are optimized to differentiate anti-ssDNA Ig from background (non-antibody) signal with mouse serum samples diluted 1:100.

## KIT CONTENTS

The microtiter well plate and all other reagents, if unopened, are stable at 2-8°C until the expiration date printed on the box label. Stabilities of the working solutions are indicated under Reagent Preparation.

**To Be Reconstituted:** Store as indicated.

Component	Preparation Instructions
<b>Wash Solution Concentrate (100x)</b> Cat. No. WB-100, 10ml	Dilute the entire volume 10ml + 990ml with distilled or deionized water into a clean stock bottle. Label as <b>Working Wash Solution</b> and store at ambient temperature until kit is used entirely.
<b>Sample Diluent Concentrate (20x)</b> Cat. No. SD-20T, 10ml	Dilute the entire volume, 10ml + 190ml with distilled or deionized water into a clean stock bottle. Label as <b>Working Sample Diluent</b> and store at 2-8°C until the kit lot expires or is used up.
<b>Anti-Mouse Ig - HRP Conjugate Concentrate (100x)</b> Part: H-MsGAM.612, 0.15ml	Peroxidase conjugated anti-Mouse Ig in buffer with protein, detergents and antimicrobial as stabilizers. Dilute fresh as needed; 10ul of concentrate to 1ml of <b>Working Sample Diluent</b> is sufficient for 1 8-well strip. Use within the working day and discard. Return 100X to 2-8°C storage.

**Ready For Use:** Store as indicated on labels.


Component	Part	Amt	Contents
<b>ssDNA Microwell Strip Plate</b>	5301	8-well strips (12)	Coated with ssDNA, and post-coated with stabilizers.
<b>Anti-DNA Calibrators</b>			
50 U/ml	51d53sB	0.65 ml	Four (4) vials, each containing anti-DNA IgG levels in arbitrary activity Units; diluted in buffer with protein, detergents and antimicrobial as stabilizers.
200 U/ml	51d53sC	0.65 ml	
500 U/ml	51d53sD	0.65 ml	
1000 U/ml	51d53sE	0.65 ml	
<b>Low NSB Sample Diluent</b>	TBTm	30 ml	Buffer with protein, detergents and antimicrobial as stabilizers. Use as is for sample dilution. See <b>Assay Design</b> , page 3. <b>Not</b> for HRP Conjugate dilution.
<b>TMB Substrate</b>	80091	12 ml	Chromogenic substrate for HRP containing TMB and peroxide.
<b>Stop Solution</b>	80101	12 ml	Dilute sulfuric acid.

### Materials Required But Not Provided:

- Pipettors and pipettes that deliver 100ul and 1-10ml. A multi-channel pipettor is recommended.
- Disposable glass or plastic 5-15ml tubes for diluting samples and Anti-Mouse Ig HRP Concentrate.
- Graduated cylinder to dilute Wash Concentrate; 0.2 to 1L.
- Stock bottle to store diluted Wash Solution; 0.2 to 1L.
- Distilled or deionized water to dilute reagent concentrates.
- Microwell plate reader at 450 nm wavelength.

## ASSAY DESIGN AND SET-UP

The **Low NSB Sample Diluent**, TBTm, lowers NSB even more than does the 1xSD20T Diluent, without diminishing true positive antibody signals, thus offering a greater discrimination between positives and non-immune samples.



### Sample Collection and Handling

Culture medium, serum and other biological fluids may be used as samples with proper dilution to avoid solution matrix interference. For **serum**, collect blood by venipuncture, allow clotting, and separate the serum by centrifugation at room temperature. For other samples, including **tissue culture media**, clarify the sample by centrifugation and/or filtration prior to dilution in Sample Diluent.

### Antibody Stability

Initial dilution of serum into **Working Sample Diluent** (WSD) is recommended to stabilize antibody activity. This enhances reproducible sampling, and stabilizes the antibody activity for years, stored refrigerated or frozen. Further dilution into **Low NSB Sample Diluent** (LNSD), which provides the lowest assay background, should be at least 10 times the initial dilution and performed the same day as the assay.

Example: Initial (1/5): 10ul serum + 40ul WSD [or 0.1ml + 0.4ml]  
Further (1/50): 10ul initial (1/5) + 90ul LNSD (1/50)

### Assay Design

Review Calculation of Results (p5-7) and Limits of the Assay (above) before proceeding:

- Select the proper sample dilutions.** Account for expected potency of positives and minimize non-specific binding (NSB) and other matrix effects; for example, non-immune samples should give net signal <0.5 OD. This is usually 1/100 or greater dilution for mouse sera with normal levels of IgG and IgM. Dilute samples in **Working Sample Diluent (1xSD20T)** or in **Low NSB Sample Diluent (TBTm)** (see above). Note: **all samples** must be diluted in the same diluent for proper comparison – either TBTm or 1xSD20T.
- Run a Sample Diluent **Blank**. This signal is an indicator of proper assay performance, especially of washing efficacy, and is used for net OD calculations, if required. Blank OD should be <0.3. **See Method A and B.**
- Run a set of Calibrators. Calibrators validate that the assay was performed to specifications, and can be used to normalize between-assay variation for enhanced precision. Reading values off a Calibrator curve, **Method C**, has limitations. See Limits of the Assay above.
- Run a range of sample dilutions for expected higher positives that allows calculation of antibody **Titer** (when specific titer is at least 4-fold higher than non-immune). **See Method D.**
- Run samples in duplicate if used for quantitation; non-immunes that are significantly lower than immunes may be run in singlicate. The Calibrators that are used for quantitation, e.g., for between-assay normalization, should be run in duplicate. When determining titer from a dilution curve, singlicates can be run if more than two dilution points are used for titer calculations.

## Plate Set-up

Bring all reagents to room temperature (18-30° C) equilibration (at least 30 minutes).

- Determine the number of wells for the assay run. Duplicates are recommended, including 8 Calibrator wells and 2 wells for each sample and control to be assayed.
- Remove the appropriate number of microwell strips from the pouch and return unused strips to the pouch. Reseal the pouch and store refrigerated.
- Add 200-300ul Working Wash Solution to each well and let stand for about 5 minutes. Aspirate or dump the liquid and pat dry on a paper towel before sample addition.

## Assay Procedure

**ALL STEPS ARE PERFORMED AT ROOM TEMPERATURE.** After each reagent addition, gently tap the plate to mix the well contents prior to beginning incubation.

### 1. 1<sup>st</sup> Incubation [100ul – 60 min; 4 washes]

- Add 100ul of calibrators, samples and controls each to pre-determined wells.
- Tap the plate gently to mix reagents and incubate for 60 minutes.
- Wash wells 4 times and pat dry on fresh paper towels. As an alternative, an automatic plate washer may be used. Improper washes may lead to falsely elevated signals and poor reproducibility.

### 2. 2<sup>nd</sup> Incubation [100ul – 30 min; 5 washes]

- Add 100ul of diluted Anti-Mouse Ig HRP to each well.
- Incubate for 30 minutes.
- Wash wells 5 times as in step 2.

### 3. Substrate Incubation [100ul – 15 min]

- Add 100ul TMB Substrate to each well. The liquid in the wells will begin to turn blue.
- Incubate for 15 minutes in the dark, e.g., place in a drawer or closet.

Note: If your microplate reader does not register optical density (OD) above 2.0, incubate for less time, or read OD at 405-410 nm (results are valid).

### 4. Stop Step [Stop: 100ul]

- Add 100ul of Stop Solution to each well.
- Tap gently to mix. The enzyme reaction will stop; liquid in the wells will turn yellow.

### 5. Absorbance Reading

- Use any commercially available microplate reader capable of reading at 450nm wavelength. Use a program suitable for obtaining OD readings, and data calculations if available.
- Read absorbance of the entire plate at 450nm using a single wavelength within 30 minutes after Stop Solution addition. If available, program to subtract OD at 630nm to normalize well background.

**LIMITS OF THE ASSAY**

**Quantitation of Antibody in a Sample**

The ELISA measures anti-ssDNA activity, a combination of antibody concentration and avidity for the ssDNA antigens. Antibodies with substantially different Ig concentrations may display similar anti-ssDNA activities, due to differences in avidity. The quantitation or activity of the samples is, therefore, appropriately expressed in activity Units (titer), rather than mass units of Ig (e.g., ug/ml).

**Calibrator Curve Quantitation**

To quantitate antibody activity from a calibrator curve (such as provided with the kit, **Method C**, the dilution curve of the samples must be parallel to the calibrator curve, to avoid different values being obtained from different regions of the curve. Antibodies that are not matched in anti-ssDNA avidity will often have non-parallel dilution curves. In these cases, antibody activity is best expressed as **Method A**: ELISA value & dilution; **Method B**: a multiple of a Positive Index titer; or **Method D**: anti-ssDNA titer from a sample dilution curve (see Calculation of Results).

**INTERPRETATION OF RESULTS**

**Calculation of Results**

Consider several data reduction methods to best represent the relationships among experimental and control groups, to determine **Positive Immune** and **Negative Non-immune**, and to **Quantitate** positive antibody levels.

**Method A. Antibody Activity [ELISA Signal & Sample Dilution]**

Represent data as net OD units (A450 signal; blank subtracted) ÷ dilution = **Total Activity Units**.

A Calibrator value in the mid-OD range (e.g., 200 U/ml) can be used to normalize inter-assay values.

**Method B. Positive Index**

Experimental sample values may be expressed relative to the values of Control or Non-immune samples, by calculation of a **Positive Index**. One typical method is as follows:

1. Calculate the net OD mean + 2 SD of the Control/Non-immune samples = **Positive Index**.
2. Divide each sample net OD by the Positive Index. Values above 1.0 are a measure of **Positive Antibody Activity**; below 1.0 are **Negative** for antibody.

A sample value would be **Positive** if significantly above the value of the pre-immune serum sample or a suitably determined non-immune panel or pool of samples, tested at the same sample dilution. This calculation **quantifies** the positive Antibody Activity level.

**Example:**

Sample	Assay Net OD		Calculated Antibody Activity	
	Control	Exptl	Control	Exptl
1	0.243	2.358	0.49	<b>4.79</b>
2	0.351	0.597	0.71	<b>1.21</b>
3	0.286	1.421	0.58	<b>2.89</b>
4	0.357	1.268	0.73	<b>2.58</b>
5	0.512	0.857	<b>1.04</b>	<b>1.74</b>
6	0.342	1.296	0.70	<b>2.63</b>
7	0.298	0.608	0.61	<b>1.24</b>
8	0.285	0.369	0.58	0.75
9	0.157	0.864	0.32	<b>1.76</b>
10	0.187	0.543	0.38	<b>1.10</b>
Mean	0.302			
SD	0.095			
Mean +2 SD	<b>0.492</b>			<b>= Positive Index</b>

**PRECAUTIONS AND SAFETY INSTRUCTIONS**

Calibrators, Sample Diluent, and Antibody HRP contain bromonitrodioxane (BND: 0.05%, w/v). Stop Solution contains dilute sulfuric acid. Follow good laboratory practices, and avoid ingestion or contact of any reagent with skin, eyes or mucous membranes. All reagents may be disposed of down a drain with copious amounts of water. MSDS for TMB, sulfuric acid and BND can be requested or obtained from the ADI website: [http://4adi.com/commerce/info/showpage.jsp?page\\_id=1060&category\\_id=2430&visit=10](http://4adi.com/commerce/info/showpage.jsp?page_id=1060&category_id=2430&visit=10)

**CALCULATION OF RESULTS (continued)**


**Method C. Use of a Calibrator Curve**

When the dilution curves of samples are parallel to the Calibrator curve (see Limits of the Assay), the anti-ssDNA activity units may be determined by interpolation from the Calibrator curve. The results may be calculated using any immunoassay software package. If software is not available, anti-ssDNA activity concentrations may be determined as follows:

1. Calculate the mean OD of duplicate samples.
2. On graph paper plot the mean OD of the calibrators (y-axis) against the concentration (U/ml) of anti-ssDNA (x-axis). Draw the best fit curve through these points to construct the calibrator curve. A point-to-point construction is most common and reliable.
3. The anti-ssDNA activity concentrations in unknown samples and controls can be determined by interpolation from the calibrator curve.
4. Multiply the values obtained for the samples by the dilution factor of each sample.
5. Samples producing signals higher than the 1000 U/ml calibrator should be further diluted and re-assayed.

Wells	Typical Results: Calibrators	A450 nm
A1,2	Negative Diluent Blank	0.12
B1,2	50 U/ml Calibrator	0.67
C1,2	200 U/ml Calibrator	1.53
D1,2	500 U/ml Calibrator	2.34
E1,2	1000 U/ml Calibrator	2.75
F1,2	Sample 1:100	2.06

Sample Result: 344 U/ml x 100 dilution = 34.4 kU/ml



**Calibrator Values**

The calibrators are dilutions of antibody reactive to ssDNA. Values are assigned as arbitrary anti-DNA activity units (see Limits of the Assay).

**CALCULATION OF RESULTS (continued)**

**Method D. Titers from Sample Dilution Curves**

The titer of antibody activity calculated from a dilution curve of each sample is recommended as the most accurate quantitative method. Best precision can be obtained using the following guidelines:


1. Use an OD value Index in the mid-range of the assay (2.0 – 0.5 OD); this provides the best sensitivity and reproducibility for comparing experimental groups and replicates. An arbitrary 1.0 OD is commonly used.
2. Prepare serial dilutions of each sample to provide a series that will produce signals higher and lower than the selected index. With accurate diluting, duplicates may not be required if at least 4 dilutions are run per sample.
3. A 5-fold dilution scheme is useful to efficiently cover a wide range which produces ODs both above and below 1.0 OD. The dilution scheme can be tightened to 3-fold or 2-fold for more precise comparative data.
4. A Calibrator value in the mid-OD range (e.g., 200 U/ml) can be used to normalize inter-assay values.

**Calculations**

1. On a log scale of inverse of Sample Dilution as the x-axis, plot the OD values of the two dilutions of each positive sample having ODs above and below the OD value of the Index (arbitrary or selected Calibrator).
2. From a point-to-point line drawn between the two sample ODs, read the dilution value (x-axis) corresponding to the OD of the selected Index = **Total Ig Antibody Activity Units**

**Example:**

II. A 1.0 OD Index was used to determine titer of 4 antibodies.



**Titer Values**

- Sample A = 1.72 kU
- Sample B = 5.70 kU
- Sample C = 1.85 kU
- Sample D = 7.90 kU

**Mouse Anti-ssDNA Antibodies Total Ig's ELISA Kit**

Cat. No. 5310, 96 tests

For Quantitation of Total anti-ssDNA Ig (IgG+IgA+IgM) in Serum, plasma or other biological fluids

For research use only (RUO), not for diagnosis, cure or prevention of the disease.



**ALPHA DIAGNOSTIC INTERNATIONAL**

6203 Woodlake Center Drive • San Antonio • Texas 78244 • USA.  
Phone (210) 561-9515 • Fax (210) 561-9544  
Toll Free (800) 786-5777

Email: [service@4adi.com](mailto:service@4adi.com)  
[www.4adi.com](http://www.4adi.com)

