

INTENDED USE

The **Horse Anti-Zaire Ebola Virus Glycoprotein (ZEBOV GP) IgG ELISA Kit** is an immunoassay suitable for quantifying IgG antibody activity specific for Zaire Ebola glycoprotein in serum or plasma of vaccinated, immunized and/or infected hosts.


This immunoassay is suitable for:

- Determining **immune status** relative to non-immune controls;
- Assessing efficacy of **vaccines**, including dosage, adjuvantcy, route of immunization and timing;
- Qualifying and standardizing vaccine batches & protocols

The assay is for research use only (RUO) and is not intended nor validated for diagnosing Ebola virus disease (EVD). Reagents contain no virus or virus –extracted-antigens.

GENERAL INFORMATION

Ebola virus (**EBOV**) causes severe disease in humans and in nonhuman primates in the form of viral hemorrhagic fever. Ebola Zaire attacks every organ and tissue in the human body except skeletal muscle and bone. Strains of Ebola include: Zaire, Sudan, Reston and Tai. All cause illness in sub-human primates. Ebola Reston has not caused illness in humans. The mortality rate of Ebola victims is between 60-90%.



Ebola virions consist of seven structural proteins.. At the center is the helical ribonucleocapsid, which consists of the genomic RNA wrapped around a polymer of nucleoproteins (**NP**). Associated with the

ribonucleoprotein is the RNA-dependent RNA polymerase (L) with the polymerase cofactor (**VP35**) and a transcription activator (**VP30**). The ribonucleoprotein is embedded in a matrix, formed by the major (**VP40**) and minor (**VP24**) matrix proteins. These particles are surrounded by a lipid membrane derived from the host cell membrane. The membrane anchors a glycoprotein (**GP1,2**) that projects 7 to 10 nm spikes away from its surface. Viral proteins (NP, GP, and VP40) are highly immunogenic and could individually or together constitute effective vaccines. Promising vaccine candidates include DNA vaccines or those based on adenoviruses, vesicular stomatitis Indiana virus (VSV) or filovirus-like particles (VLPs) as all of these candidates could protect nonhuman primates from Ebola virus-induced disease.

PRINCIPLE OF THE TEST

The Anti-ZEBOV IgG/IgM ELISA kits are based on the binding of antibodies (IgG/IgM) in samples to the recombinant, purified ZEBOV antigen (GP, NP or VP40) immobilized on the microwells. Bound antibody is detected by anti-Horse IgG or IgM-HRP conjugate. After a washing step, chromogenic substrate (TMB) is added and color (blue) is developed, which is directly proportional to the amount of anti-ZEBOV IgG or IgM present in the sample. Stop Solution is added to terminate the reaction (convert blue to yellow color), and A450 is then measured using an ELISA reader at 450nm. The presence of antibody (IgG/IgM) in samples is determined relative to anti-ZEBOV IgG/IgM Calibrators and Controls.

KIT CONTENTS

The microtiter well plate and all other reagents, if unopened, are stable at 2-8°C until the expiration date printed on the box label. Stabilities of the working solutions are indicated under Reagent Preparation.

To Be Reconstituted: Store as indicated.

Component	Preparation Instructions
Wash Solution Concentrate (100x) Cat. No. WB-100, 10ml	Dilute the entire volume 10ml + 990ml with distilled or deionized water into a clean stock bottle. Label as Working Wash Solution and store at 4°C for long term and ambient temp. for short term.
Sample Diluent Concentrate (20x) Cat. No. SD-20T, 10ml	Dilute the entire volume, 10ml + 190ml with distilled or deionized water into a clean stock bottle. Label as Working Sample/Conjugate Diluent and store at 2-8°C until the kit lot expires or is used up.
Anti-Horse IgG-HRP Conjugate Concentrate (100x) Part: H-HgG.211, 0.15ml	Peroxidase conjugated anti-Horse IgG in buffer with detergents and antimicrobial as stabilizers. Dilute fresh as needed; 10ul of concentrate to 1ml of Working Sample/Conjugate Diluent is sufficient for 1 8-well strip. Use within the working day and discard. Return 100X to 2-8°C storage.

Ready For Use: Store as indicated on labels.

Component	Part	Amt	Contents
ZEBOV GP Coated Strip Plate	320621	8-well strips (12)	Coated with purified recombinant ZEBOV GP, and post-coated with stabilizers.
Anti-ZEBOV GP Calibrators			
1 U/ml	320622B	0.65 ml	Four (4) vials, each containing anti-ZEBOV GP; in buffer with antimicrobial as stabilizers.
2.5 U/ml	320622C	0.65 ml	
5 U/ml	320622D	0.65 ml	
10 U/ml	320622E	0.65 ml	
Anti-x-ZEBOV GP IgG Positive Control	320670-PC	0.65 ml	Horse serum with anti-ZEBOV GP reactivity; Net OD > 0.5
Low NSB Sample Diluent Reduces non-specific binding	TBTm Not for HRP Conjugate dilution.	30 ml	Buffer with protein, detergents and antimicrobial as stabilizers. Use as is for sample dilution. See Assay Design , page 3.
TMB Substrate	80091	12 ml	Chromogenic substrate for HRP containing TMB and peroxide.
Stop Solution	80101	12 ml	Dilute sulfuric acid.

Materials Required But Not Provided:

- Pipettors and pipettes that deliver 100ul and 1-10ml.
- Disposable glass or plastic 5-15ml tubes for diluting samples and Anti-Horse IgG HRP Concentrate.
- Stock bottle to store diluted Wash Solution; 0.2 to 1L.
- Distilled or deionized water to dilute reagent concentrates.
- Microwell plate reader at 450 nm wavelength and ELISA plate washer

ASSAY DESIGN AND SET-UP

Sample Collection and Handling

Serum and other biological fluids may be used as samples with proper dilution to avoid solution matrix interference. For **serum**, collect blood by venipuncture, allow clotting, and separate the serum by centrifugation at room temperature. If samples will not be assayed immediately, store refrigerated for up to a few weeks, or frozen for long-term storage.

Sample Dilution & Antibody Stability

Prepare an initial sample dilution (1:10 or 20 ul sample into 180 ul) of **Working Sample Diluent** in order to stabilize antibody activity. This enhances reproducible sampling, and stabilizes the antibody activity for months, stored refrigerated or frozen. Additional dilution (1:10 of the initial stock for a final dilution of 1:200) into **Low NSB Sample Diluent** provides low assay background and good discrimination of specific signal. It is possible to change the testing dilution to 1:50-1:500 depending upon the actual sample background. All sample dilutions in Low NSB should be at least 5 times the initial dilution and performed the same day as the assay. Do not store test dilutions.

Assay Design

Review Interpretation of Results (p5-7) before proceeding:

- Select the proper sample dilutions accounting for expected potency of positives and minimizing non-specific binding (NSB) and other matrix effects; for example, net signal for non-immune samples should be lower than the **1 U/ml Calibrator**. This is usually 1:200 or greater dilution for Horse serum with normal levels of IgG and IgM.
- Run the Horse Anti-ZEBOV GP IgG Positive Control; net OD > **0.5**.
- Run a Sample Diluent **Blank**. This signal is an indicator of proper assay performance, especially of washing efficacy, and is used for net OD calculations, if required. Blank OD should be <0.3.
- Run a set of **Calibrators**, which validate that the assay was performed to specifications: **10 U/ml** should give a high signal (>1.5 OD); **1 U/ml** should give a low signal which can be used to discriminate at the Positive/Negative threshold (see Interpretation of Results, p. 5).

Plate Set-up

Bring all reagents to room temperature (18-30° C) equilibration (at least 30 minutes).

- Determine the number of wells for the assay run. Duplicates are recommended, including 8 Calibrator wells and 2 wells for each sample control to be assayed.
- Remove the appropriate number of microwell strips from the pouch and return unused strips to the pouch. Reseal the pouch and store refrigerated.
- Add 200-300ul Working Wash Solution to each well and let stand for about 5 minutes. Aspirate or dump the liquid and pat dry on a paper towel before sample addition.

Assay Procedure

ALL STEPS ARE PERFORMED AT ROOM TEMPERATURE. After each reagent addition, gently tap the plate to mix the well contents prior to beginning incubation.

1. 1st Incubation [100ul – 60 min; 4 washes]

- Add 100ul of calibrators, samples and controls each to pre-determined wells.
- Tap the plate gently to mix reagents and incubate for 60 minutes.
- Wash wells 4 times and pat dry on fresh paper towels. As an alternative, an automatic plate washer may be used. Improper washes may lead to falsely elevated signals and poor reproducibility.

2. 2nd Incubation [100ul – 30 min; 5 washes]

- Add 100ul of diluted Anti-Horse IgG HRP to each well.
- Incubate for 30 minutes.
- Wash wells 5 times as in step 2.

3. Substrate Incubation [100ul – 15 min]

- Add 100ul TMB Substrate to each well. The liquid in the wells will begin to turn blue.
- Incubate for 15 minutes in the dark, e.g., place in a drawer or closet.

Note: If your microplate reader does not register optical density (OD) above 2.0, incubate for less time, or read OD at 405-410 nm (results are valid).

4. Stop Step [Stop: 100ul]

- Add 100ul of Stop Solution to each well.
- Tap gently to mix. The enzyme reaction will stop; liquid in the wells will turn yellow.

5. Absorbance Reading

- Use any commercially available microplate reader capable of reading at 450nm wavelength. Use a program suitable for obtaining OD readings, and data calculations if available.
- Read absorbance of the entire plate at 450nm using a single wavelength within 30 minutes after Stop Solution addition. If available, program to subtract OD at 630nm to normalize well background.

PRECAUTIONS AND SAFETY INSTRUCTIONS


Calibrators, Sample Diluent, and Antibody HRP contain bromonitrodioxane (BND: 0.05%, w/v). Stop Solution contains dilute sulfuric acid. Follow good laboratory practices, and avoid ingestion or contact of any reagent with skin, eyes or mucous membranes. All reagents may be disposed of down a drain with copious amounts of water. MSDS for TMB, sulfuric acid and BND can be requested or obtained from the ADI website: http://4adi.com/commerce/info/showpage.jsp?page_id=1060&ca tegory_id=2430&visit=10

INTERPRETATION OF RESULTS

Method A. Antibody Activity Threshold Index

Compare Samples to 1 U/ml Calibrator or Internal Control
= Positive/Negative Cut-off.

Example:



Results

The **sensitivity** of the assay to detect anti-GP IgG, from either natural infection or vaccination, is controlled so that the 1 U/ml Calibrator represents a threshold OD for most true positives in Horse serum diluted to 1:200 or greater. Visual inspection of the data in the above graph shows the following:

Calibrators – dilution curve of an anti-ZEBOV GP antibody, derived from GP immunization, shows the OD range of the assay; high value indicates optimal sensitivity of the assay.

1 U/ml: a 'Cut-off' line has been drawn to indicate a threshold distinguishing between **Positive/Negative**. The is not a clear-cut threshold, rather a low OD area that could represent either low positives or high background negatives.

Positive Control – Horse serum showing natural reactivity to ZEBOV GP; net OD > 0.5. This Control can be used to normalize between-assay variation.

Internal Control – a true positive from an immune Horse that represents the investigator's experience in distinguishing low positive from negative samples (not in kit). This should be run in each assay to supplement the 1 U/ml Calibrator for Positive/Negative discrimination purposes.

ASSAY PERFORMANCE

Samples A,B,C,D – 3 samples (A, B, C) are **negative**: below the threshold; 1 sample (D) is **positive**: clearly above the threshold.

The 1 U/ml Calibrator can be used to calculate a **Threshold Index** that numerically discriminates Positive/Negative (see p6):

- ❖ Divide each Sample net OD by the 1 U/ml Calibrator net OD. Values above 1.0 are a measure of **Positive** Antibody Activity; below 1.0 are **Negative** for antibody.

Example:

Horse Serum IgG

A panel of non-ZEBOV-immunized Horse sera was tested for anti-ZEBOV GP IgG (1:200 dilution). **Threshold Index** was calculated using the 1 U/ml Cal.

Results

Anti-ZEBOV GP IgG: two (2) samples, including horse, were positive (above the 1.0 threshold) at 1:200 dilution; the other twelve (12) samples were negative (below 1.0 threshold).

Notes:

- Positives** may be due to prior encounter with the virus or non-Ebola proteins with common epitopes, or may be an aspect of the innate immune repertoire. All of the dogs and wolves in the panel had been immunized with rabies vaccine.
- The **sensitivity** of the assay may be adjusted by changing the sample dilutions: a) increase dilution (e.g., 1/500) to lower the signals of borderline positives to negative; b) decrease dilution (e.g., 1/50) to convert borderline samples to positive. With the latter, the values of negatives may increase, so an alternative threshold should be considered using known negatives to develop a **Positive Index** (see below) or use an **Internal Control** (Page 5).
- Samples with Threshold Index values approaching 5.0 should be quantified using the titer Method C (page 7).

Additional testing of human, mouse and pig data is available on the website.

http://www.4adi.com/objects/catalog/product/extras/Ebola_Marburg_Vaccines_ELISA_Flr.pdf

B. Positive Index

Experimental sample values may be expressed relative to the values of Control or Non-immune samples, by calculation of a **Positive Index**. One typical method is as follows:

- Calculate the net OD mean + 2 SD of the Control/Non-immune samples = **Positive Index**.
- Divide each sample net OD by the Positive Index. Values above 1.0 are a measure of **Positive** Antibody Activity; below 1.0 are **Negative** for antibody.

INTERPRETATION OF RESULTS (cont)

A sample value would be **Positive** if significantly above the value of the pre-immune serum sample or a suitably determined non-immune panel or pool of samples, tested at the same sample dilution.

This calculation also **quantifies** the positive Antibody Activity level, assigning a higher value to samples with higher Antibody Activity, and vice versa.

Method C. Titers from Sample Dilution Curves

The titer of elevated antibody activity calculated from a dilution curve of each sample is recommended as the most accurate quantitative method. Best precision can be obtained using the following guidelines:

- Use an OD value Index in the mid-range of the assay (2.0 – 0.5 OD); this provides the best sensitivity and reproducibility for comparing experimental groups and replicates. An arbitrary 1.0 OD is commonly used.
- Prepare serial dilutions of each sample to provide a series that will produce signals higher and lower than the selected index. With accurate diluting, duplicates may not be required if at least 4 dilutions are run per sample.
- A 5-fold dilution scheme is useful to efficiently cover a wide range which produces ODs both above and below 1.0 OD. The dilution scheme can be tightened to 3-fold or 2-fold for more precise comparative data.
- The Positive and Sensitivity Control values can be used to normalize inter-assay values.

Calculations

- On a log scale of inverse of Sample Dilution as the x-axis, plot the OD values of the two dilutions of each positive sample having ODs above and below the OD value of the Index (arbitrary or selected Calibrator).
- From a point-to-point line drawn between the two sample ODs, read the dilution value (x-axis) corresponding to the OD of the selected Index
= **IgG Antibody Activity Units**

PRODUCT SPECIFICATIONS

Specificity

The following are used for coating antigens:

Ebola Nucleoprotein (NP) antigen: Zaire Ebola NP, full length, His-tag, E. coli expressed protein (>95% pure).

Ebola VP40 antigen: Zaire Ebola VP40, full length, no tag, E. coli expressed protein (>95% pure).

Ebola Glycoprotein (GP) antigen: Zaire Ebola GP, full length minus TM domain, His-tag, Sf9 expressed protein (>95% pure).

Sequence Conservation in Ebola subtypes.

	Zaire	Bundi-bugyo	Cote D'Ivo.	Reston	Sudan	Marburg
NP	100%	75%	69%	69%	67%	53%
VP40	100%	82%	69%	74%	75%	34%
GP	100%	65%	64%	57%	54%	NA

Although the 3 Ebola virus antigens are significantly conserved in various Ebola serotype and also in related Marburg viruses, it is not known if the antibodies elicited by one Ebola strain will be cross-reactive with antigens from other strains.

Alpha Diagnostic Intl (www.4adi.com) 320690/160721A Page 7

Instruction Manual No. M-AE-320690-1

Recombivirus™ Horse Anti-Zaire Ebola Virus Glycoprotein (ZEBOV GP) IgG ELISA Kit

ELISA KIT # AE-320690-1, 96 tests

For the Quantitation of Anti-ZEBOV GP IgG in Horse Serum or Plasma or other biological fluids

For research use only, not for diagnostic or therapeutic use.



**ALPHA DIAGNOSTIC
INTERNATIONAL**

6203 Woodlake Center Drive • San Antonio • Texas 78244 • USA.

Phone (210) 561-9515 • Fax (210) 561-9544
Toll Free (800) 786-5777 Email: service@4adi.com

Draft version-Please consult the manual supplied with the kit for any lot specific change.